



Respironics Inc.
1001 Murry Ridge Lane
Murrysville, PA 15668 USA



1078520
JAW 11/09/2010
© 2010 Koninklijke Philips Electronics N.V. All rights are reserved.



GoLife | INSTRUCTIONS FOR USE

PHILIPS
RESPIRONICS

Table of Contents

2	Package Contents
3	Intended Use
3	Warnings
7	Symbols
8	Before Use
8	Cleaning Instructions
10	System One Resistance Control
11	Attaching the Cushion
11	Removing the Cushion
12	Putting on the Mask and Achieving the Right Fit
14	Removing the Mask
14	Removable Elbow with Integrated Tube
15	Tube Management
16	Specifications
17	Disposal
17	Storage Conditions
18	Reorder Information
19	Limited Warranty

Package Contents

- GoLife Nasal Pillows Mask with Cushions
- Headgear and Headgear Frame
- Instructions for Use
- QuickStart Fitting Tips



Note:

Not all packages contain listed items. Please contact your Home Care Provider or Philips Respironics for additional items.







Figure 1

- A = Top Headgear Strap
- B = Side Headgear Strap
- C = Headgear Frame
- D = Cushion Nub
- E = Swivel (Integrated with the 15mm Tubing)
- F = Tube Management Clip
- G = Mask Elbow with Integrated 15mm Tubing
- H = Built-in Exhalation (Do not block)
- I = Pillows Cushion

GoLife

Intended Use

The GoLife Nasal Pillows Mask is intended to provide an interface for CPAP or bi-level therapy. This mask is intended for single-patient use in the home or multi-patient use in the hospital/institutional environment. This mask is to be used on patients greater than 66 lbs /30 kg.

-  **Important Note:** An exhalation port is built into the GoLife Nasal Pillows Mask so a separate exhalation port is not required.
-  **Note:** This mask does not contain natural rubber latex.
-  **Note:** This mask does not contain bis (2-ethylhexyl) phthalate (DEHP).

Caution

U.S. Federal law restricts this device to sale by or on the order of a physician.

Warnings

- This mask is not suitable for providing life support ventilation.
- This mask is designed for use with CPAP or bi-level systems recommended by your health care professional or respiratory therapist. Do not wear this mask unless the CPAP or bi-level system is turned on and operating properly. **Do not block or try to seal the exhalation port. Explanation of the Warning:** CPAP systems are intended to be used with special masks with connectors which have vent holes to allow continuous flow of air out of the mask. When the

Warnings (continued)

CPAP machine is turned on and functioning properly, new air from the CPAP machine flushes the exhaled air out through the attached mask exhalation port. However, when the CPAP machine is not operating, enough fresh air will not be provided through the mask, and exhaled air may be rebreathed. This warning applies to most models of CPAP systems. Rebreathing of exhaled air for longer than several minutes can, in some circumstances, lead to suffocation.

- If oxygen is used with the device, the oxygen flow must be turned off when the device is not operating. **Explanation of the Warning:** When the device is not in operation, and the oxygen flow is left on, oxygen delivered into the ventilator tubing may accumulate within the device enclosure. Oxygen accumulated in the device enclosure will create a risk of fire.
- Oxygen supports combustion. Oxygen should not be used while smoking or in the presence of an open flame.



Warning

A warning indicates the possibility of injury to the user or operator.



Caution

A caution indicates the possibility of damage to the device.



Note

A note places an emphasis on an operating characteristic.

Warnings (continued)






- At a fixed flow rate of supplemental oxygen flow, the inhaled oxygen concentration will vary, depending on the pressure settings, patient breathing pattern, mask selection, and the leak rate. This warning applies to most types of CPAP and bi-level machines.
- Some users may experience skin redness, irritation, or discomfort. If this happens, discontinue use and contact your healthcare professional.
- At low CPAP or EPAP pressures, the flow through the exhalation port may be inadequate to clear all exhaled gas from the tubing. Some rebreathing may occur.
- If an additional exhalation device is added to the patient circuit, you may need to adjust the pressure level to compensate for the additional leak of the exhalation device.
- The GoLife mask has been validated for use with Philips Respironics autotitrating CPAP and bi-level therapy devices that have the System One resistance compensation feature and the Philips Respironics Trilogy. The System One feature on the CPAP or bi-level device must be enabled by the provider in accordance with the therapy device manual.
- Disconnect sensing systems on ventilators may not work as expected with GoLife pillows. Caregiver should visually monitor patients for disconnects or use an alternate interface. Do not rely on any single

Warnings (continued)

alarm to detect a patient disconnect condition. The Low Tidal Volume, Low Minute Ventilation, Low Respiratory Rate and Apnea alarms (if available) should be used in conjunction with this mask to ensure proper ventilation. **Explanation of the Warning:** The ventilator disconnect alarm attempts to detect changes in patient circuit resistance (including the filter, humidifier, tubing, and mask) while connected to the patient. In the event the patient becomes disconnected from the interface during ventilation with the GoLife pillows, the circuit resistance change is not sufficient to activate the disconnect alarm. Therefore, the circuit disconnect alarm may not function.

- Use of a nasal or full face mask may cause tooth, gum, or jaw soreness or aggravate an existing dental condition. Consult your physician or dentist if symptoms occur.
- Hand wash prior to first use. Inspect the mask for damage or wear (cracking, crazing, tears, etc). Discard and replace any components as necessary.

Symbols

Symbol	Description	Symbol	Description
	Warning or Caution		Does Not Contain Natural Rubber Latex
	Note		System One Resistance Control
	Consult Instructions for Use		

Before Use

- Read and understand the instructions completely.
- Hand wash the mask.
- Wash your face.
- Remove the QuickStart fitting tips (if attached to the headgear).

Cleaning Instructions

Hand wash the entire mask, including headgear frame and straps before first use. The pillows cushions and tubing should be hand washed daily and the headgear frame and straps should be hand washed weekly, or as needed. The headgear frame and straps do not need to be removed for daily cleaning.

1. Hand wash the mask and headgear straps in warm water with a mild liquid dish washing detergent.

⚠Caution: Do not use bleach, alcohol, cleaning solutions containing bleach or alcohol, or cleaning solutions containing conditioners or moisturizers.


2. Rinse thoroughly with water and air dry out of direct sunlight. Make sure the mask is dry before use. Lay the headgear flat or line dry. Do not place the headgear into the dryer.

Cleaning Instructions (continued)

⚠ Warning: Inspect the mask for damage or wear (cracking, crazing, tears, etc). Discard and replace any components as necessary.

For **multi-patient use** in the hospital/institutional environment, use the **Mask Cleaning and Disinfection Instructions** to reprocess the mask between patients. These instructions can be obtained by visiting us online at www.philips.com/respironics or by contacting Philips Respironics Customer Service at 1-800-345-6443 (USA or Canada) or at 1-724-387-4000.

Philips Respironics System One Resistance Control

This mask uses the System One Resistance Control setting (e.g., X1). The following symbol may appear on your Philips Respironics therapy device (e.g.,  ← X1). This represents the level of mask resistance compensation. Match the setting to the mask as appropriate.

GoLife Nasal Pillows	
Size	Setting
P	X4
S	X3
M	X3
L	X2

Important Notes:

- Verify the System One setting if the mask or mask cushion changes.
- The System One setting is not compatible with masks requiring a separate/additional exhalation device.

If your provider has locked the resistance compensation setting, you can view the setting but cannot change it, and the screen will display a lock symbol. If your provider has disabled the resistance compensation, you will not see this setting.

Attaching the Cushion

1. To attach a cushion, pinch the sides of the cushion so it flexes inward (Figure 2).
2. With the cushion flexed inward, place the opening over the tubing post on the frame (Figure 3).
3. Push the ends of the cushion into the frame so the round nubs on the cushion fill the circles on the right side of the frame, and then on the left side of the frame (Figure 4). Massage the cushions into place.

Removing the Cushion

To remove a cushion, grab one side of the cushion and peel it off of the frame.




Figure 2




Figure 3




Figure 4

Putting on the Mask and Achieving the Right Fit

1. Pull the elastic headgear on over your head (Figure 5).
2. Place the cushion into your nostrils. Select a cushion that fills your nostrils comfortably without any gaps between the cushion and the walls of your nostrils (Figure 6).
3. The adjustable top strap should be placed comfortably over the top of your head so that the headgear side straps rest just above the ears, similar to where eye or sunglasses would rest. The back strap will sit at the base of your head (Figure 7).




Figure 5



Figure 6





Figure 7

Putting on the Mask and Achieving the Right Fit (continued)

4. Evenly adjust the side straps at the same time by pulling the fasteners off of the headgear and adjusting the length of the straps. Reattach the fastener when the cushion fits snugly in your nostrils (Figure 8) and the mask and headgear feel comfortable.



 **Note:** Do not overtighten the headgear side straps.

 **Figure 8**

5. Connect the tubing included with the CPAP or bi-level therapy device to the tubing swivel.
6. Lie down and turn on the airflow.
7. Make final adjustments while lying down.

Removing the Mask

Easily remove the mask by grabbing the back headgear strap and slide the headgear forward over your head.

Removable Elbow with Integrated Tube

The exhalation elbow with integrated tubing can be removed for cleaning.

1. Remove the cushion (Figure 9).
2. Hold the headgear frame with the elbow between your first and second fingers. With your thumb, push the back of the elbow through the frame (Figure 10).

⚠ Caution: Do not pull on the 15mm tubing or damage may occur.

3. Clean per instructions.
4. Reattach the clean elbow with integrated tubing to the frame (Figure 11).
5. Attach the pillows cushion.




Figure 9




Figure 10




Figure 11

Tube Management

Use the moveable tube management clip to position the CPAP or bi-level tubing to fit to your needs. The fabric ring can be moved and placed anywhere along the tube and the clip can be attached to your clothing or pillow case for easier tube management (Figure 12).

The elbow swivels so the tubing can go over your head and the tube management clip can be used to fasten the tubing to the headgear (Figure 13).




Figure 12





Figure 13

Specifications

⚠ Warning: The technical specifications of the mask are provided for your clinician to check that they are compatible with the CPAP or bi-level device. If used outside these specifications, or if used with incompatible devices, the comfort of the mask may not be effective, optimum therapy may not be achieved, and leak, or variation in the rate of leak, may affect the CPAP or bi-level device function.

Intentional Leak



GoLife

Specifications (continued)

Pressure Drop (cm H₂O)

	50 SLPM	100 SLPM
All cushion sizes	<2.9	<9.5

Deadspace Volume

All cushion sizes <27 mL

Disposal

Dispose of in accordance with local regulations.

Storage Conditions

Temperature: -4° to 140° F (-20° to +60° C)

Relative Humidity: 15% to 95% non-condensing

Reorder Information

1073114	P-S-M-L GOLIFE NSL MSK PLW W/HGR
1073328	GOLIFE W/O PLW W/O HGR
1073329	P-S-M-L GOLIFE W/O HGR
1073332	RP P GOLIFE PILLOWS CUSHION
1073333	RP S GOLIFE PILLOWS CUSHION
1073334	RP M GOLIFE PILLOWS CUSHION
1073335	RP L GOLIFE PILLOWS CUSHION
1073336	RP GOLIFE HEADGEAR
1073337	RP GOLIFE TUBE MANAGEMENT CLIP 2PK
1078407	RP GOLIFE SWIVEL TUBING (18"/45.7cm)
1078408	RP GOLIFE FRAME ONLY

Visit www.philips.com/respironics for a complete list of products and reorder information, or contact Philips Respironics Customer Service at 1-800-345-6443 (USA or Canada) or at 1-724-387-4000 for additional information.

Limited Warranty

Respironics, Inc. warrants that its mask systems (including mask frame and cushion) (the "Product") shall be free from defects of workmanship and materials for a period of ninety (90) days from the date of purchase (the "Warranty Period").

If the Product fails under normal conditions of use during the Warranty Period and the Product is returned to Respironics within the Warranty Period, Respironics will replace the Product. This warranty is non-transferable and only applies to the original owner of the Product. The foregoing replacement remedy will be the sole remedy for breach of the foregoing warranty. This warranty does not cover damage caused by accident, misuse, abuse, negligence, alteration, failure to use or maintain the Product under conditions of normal use and in accordance with the terms of the product literature, and other defects not related to materials or workmanship. This warranty does not apply to any Product that may have been repaired or altered by anyone other than Respironics.

Respironics disclaims all liability for economic loss, loss of profits, overhead, or indirect, consequential, special or incidental damages which may be claimed to arise from any sale or use of the Product. Some states do not allow the exclusion or limitation of incidental or consequential damages, so the above limitation or exclusion may not apply to you.

THIS WARRANTY IS GIVEN IN LIEU OF ALL OTHER EXPRESS WARRANTIES. IN ADDITION, ANY IMPLIED WARRANTY, INCLUDING ANY WARRANTY OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE, IS SPECIFICALLY DISCLAIMED. SOME STATES DO NOT ALLOW DISCLAIMERS OF IMPLIED WARRANTIES, SO THE ABOVE LIMITATION MAY NOT APPLY TO YOU. THIS WARRANTY GIVES YOU SPECIFIC LEGAL RIGHTS, AND YOU MAY ALSO HAVE OTHER RIGHTS, WHICH VARY FROM STATE TO STATE.

To exercise your rights under this limited warranty, contact your local authorized Respironics, Inc. dealer or Respironics, Inc. at 1001 Murry Ridge Lane, Murrysville, Pennsylvania 15668, 1-800-345-6443 (USA and Canada only) or 1-724-387-4000.

GoLife